

CLARIFICATIONS No.1

Issued on 29.12.2015.

ICB No. IOP/09-2015/NCE/2

8042/1
29.12.15

Project: Public Sector Research and Development

Subproject: New Capital Equipment

Procurement: Procurement of Capital Equipment

Qtns. No.	Reference in PN/ Tender dossier	Questions	Answers
1	Annex 1, Technical part, Technical Specifications	<p>Questions dated: 31/07/2015</p> <p>In the Technical Specifications for LOT 13- Liquid chromatography for procurement number IOP/09-2015/NCE/2, it is requested:</p> <p>Flow rate range, 05-10 ml/min, increment: 0,001 ml/min.</p> <p>Is it acceptable, according to technical instructions, page number 69, where is stated that offered goods should be „the newest and most up-to-date models and should be comprised with all recent upgrades in design and materials“</p> <p>to offer an appropriate instrument which is currently actual, with an improved newest technique for which is sufficient:</p> <p>Flow rate range, 0,001-5,000 ml/min, increment: 0,001 ml/min.</p> <p>The instrument will be used primarily for mass spectrometry where flows larger than 1 ml/min are not needed, as for all other characteristics, they accede required.</p>	<p>Yes, it is acceptable to offer Flow rate range, 0,001-5,000 ml/min, increment: 0,001 ml/min.</p>

Travis Taylor