

IZMENA Br: 3

ЈУП ИСТРАЖИВАЊЕ И РАЗВОЈ ДОО БЕОГРАД

Бр.

383
20.01 2016

Tenderske dokumentacije za nabavku Dobara

БЕОГРАД

Indetifikacioni broj: MOP br: IOP/09-2015/NCE/2

Predmet: Nabavka nove kapitalne opreme

Mesto: Republika Srbija

Sledeće izmene su načinjene u Tenderskoj dokumentaciji za nabavku Dobara:

- Originalni tekst Odeljka VII. Pregled zahteva, Tehnička specifikacija ANNEX I za Lot No.: 20 stavku 20.1 - Animal Monitoring System, for 4 rats

3. Indirect open circuit calorimeter system (VO₂ / VCO₂ / Temperature) for 4 lab animals (mouse up to dog) in the home cage. Complete and consisting of (for connecting to a PC): Air supply pump; 4 Mass flow controllers/ High speed gas drying unit; Air valve multiplexer for 4 measuring channels and 1 reference channel, / Air sample pump/ CO₂ measuring unit/ O₂ measuring unit/ control unit connecting all components.

je izmenjen sledećim:

3. Indirect open circuit calorimeter system (VO₂ / VCO₂ / Temperature) for 4 lab animals (mouse up to rats) in the home cage. Complete and consisting of (for connecting to a PC): Air supply pump; 4 Mass flow controllers/ High speed gas drying unit; Air valve multiplexer for 4 measuring channels and 1 reference channel, / Air sample pump/ CO₂ measuring unit/ O₂ measuring unit/ control unit connecting all components.

- Originalni tekst Odeljka VII. Pregled zahteva, Tehnička specifikacija ANNEX I za Lot No.: 19 stavku 19.1 - Oxygraph System

1. Main Basic
2. Peltier Temp-Control: 2 to 47 °C; ±0.001 °C
3. Electromagnetic Stirrer Twin-Control
5. Barometric Pressure Transducer
6. Assembly Kit, incl. 2 Glass Chambers
- 7 Polarographic Oxygen Sensors (POS)
8. Service Kit
9. Acquisition & Analysis Software
- 10.ISS-Integrated Suction System
- 11.Titration Set: 6 10 µl, 6 25 µl, 2 50 µl
12. Compendium, incl. USB-Stick with Manual
13. Titration - injection micro pump - module (Programmable, automatic titration regimes, with titration volumes of 0.05 to 250 µl, variable titration intervals and duration of titration pulses)
- 14.Workshop / 4 Days / Registration: waived
- 14.Delivery and installation
- 15.Delivery time: 90 days
- 16.Warranty: 1 year
- 17.Delivery place: Scientific Research Institution name: Institute for Biological Research "Sinisa Stankovic", Department of Physiology
Bul. despota Stefana 142, 11060 Belgrade

je izmenjen sledećim:

1. Main Basic
2. Peltier Temp-Control: 2 to 47 °C; ±0.001 °C
3. Electromagnetic Stirrer Twin-Control
5. Barometric Pressure Transducer
6. Assembly Kit, incl. 2 Glass Chambers
- 7 Polarographic Oxygen Sensors (POS)
8. Service Kit
9. Acquisition & Analysis Software
- 10.ISS-Integrated Suction System
- 11.Titration Set: 6 10 µl, 6 25 µl, 2 50 µl
12. Compendium, incl. USB-Stick with Manual
13. Titration - injection micro pump - module (Programmable, automatic titration regimes, with titration volumes of 0.05 to 250 µl, variable titration intervals and duration of titration pulses)
14. Fluorescence LED-2Module
- 15.Workshop / 4 Days / Registration: waived
- 16.Delivery and installation
- 17.Delivery time: 90 days
- 18.Warranty: 1 year
- 19.Delivery place: Scientific Research Institution name: Institute for Biological Research "Sinisa Stankovic", Department of Physiology
Bul. despota Stefana 142, 11060 Belgrade

- Originalni tekst Odeljka VII. Pregled zahteva, Tehnička specifikacija ANNEX I za Lot No.: 11 stavku 11.1 - D-T neutron generator

3. Delivery period : 90 days

je izmenjen sledećim:

3. Delivery period : 7 months

- Originalni tekst Odeljka VII. Pregled zahteva, Tehnička specifikacija ANNEX I za Lot No.: 3 stavku 3.1 - Germanium detector

Active area: 5000 mm²

Active thickness: 30 mm

FWHM resolution: 1950 eV at 1.33 MeV; 720 eV at 122 keV; 480 eV at 5.9 keV

Endcap diameter: 101.6 mm

Standard configuration: Vertical dipstick slimline cryostat with 30 liter dewar, 0.6 mm Carbon Epoxy window, Preamplifier, 3 meter Cable set, Detector manual

Delivery and installation

Delivery time: 90 days

Warranty period : 12 months

Delivery place: Faculty of Science, Trg Dositeja Obradovića 3, 21000 Novi Sad

je izmenjen sledećim:

Active area: minimum 5000 mm²

Active thickness: minimum 30 mm

FWHM resolution: not lower than 1950 eV at 1.33 MeV; 720 eV at 122 keV; 480 eV at 5.9 keV

Endcap diameter: minimum 101.6 mm

Standard configuration: Vertical dipstick slimline cryostat with 30 liter dewar, 0.6 mm Carbon Epoxy window, Preamplifier, 3 meter Cable set, Detector manual

Delivery and installation

Delivery time: 90 days

Warranty period : 12 months

Delivery place: Faculty of Science, Trg Dositeja Obradovića 3, 21000 Novi Sad

- Originalni tekst Odeljka III, Stručna ocen I kriterijumi kvalifikacije, 2. Kvalifikacija (UP 36), 2.3 Poslovni kapacitet ponuđača, 2.3.1 Poslovni kapacitet, glasi:

Ponuđač je uspešno dostavio robu tokom protekle 3 godine, pre roka za dostavljanje ponuda, pod jednim ugovorom slične prirode (to jest Laboratorijsku I opremu za istraživanje I razvoj) čija vrednost prelazi ili je bar jednaka vrednosti ponude (finansijski predlog).

je izmenjen sledećim:

Ponuđač je uspešno dostavio robu tokom protekle 4 godine (2012, 2013, 2014 i 2015) pre roka za dostavljanje ponuda, pod maksimalno dva ugovora slične prirode (to jest Laboratorijsku I opremu za istraživanje I razvoj) čija vrednost prelazi ili je bar jednaka vrednosti ponude (finansijski predlog).

- Odeljak IV, Konkursni obrasci briše se:

Form EXP Iskustvo (nastavak)

Molimo vas da u prilogu pogledate modifikovane Lotove 3,11,19 i 20 ANNEX-a I, Tehničke specifikacije

Svi ostali uslovi Tenderske dokumentacije ostaju nepromenjeni.



Gorenavedene izmene su sastavni deo Tenderske dokumentacije.