

CORRIGENDUM NO.1 OF THE BIDDING DOCUMENT

Published on 10.02.2022.

N/D/01/2021

A corrigendum is made of the Bidding Document for the **procurement of Equipment for the development of regional training centres and modernisation of secondary vocational schools within School Modernisation Programme, reference number N/D/01/21**, in the following part:

Bidding Document, Price Schedule form for the equipment for the Medical school „dr Andra Jovanović” in Šabac and Technical Specification form for the equipment for the Medical school „dr Andra Jovanović” in Šabac, **item 1.3 Anti decubite mattress**:

Anti decubite mattress: Product characteristics and technical data:

- Dimensions of the inflated mattress 290x185
- Can be used as a mattress topper or as a replacement mattress with the addition of 7.5cm wide foam
- Material of cells, bases: nylon +PVC
- Power supply: 220-240V 50Hz
- Atmospheric pressure - from 700hPa to 1013.25 hPa
- Cycle time : Fixed 9.6 min @50Hz
- Fuse rating – T1AL, 250V
- Weight up to 2.2 kg

The description of the item 1.3 Anti decubite mattress is changed and now reads:

Anti decubite mattress: Product characteristics and technical data:

- Dimensions of the inflated mattress 200x90x12.7 cm
- Can be used as a mattress topper or as a replacement mattress with the addition of 7.5cm wide foam
- Material of cells, bases: nylon +PVC
- Compressor dimensions 290x185x126 mm
- Power supply 220-240V 50Hz
- Atmospheric pressure from 700hPa to 1013.25 hPa
- Cycle time : Fixed 9.6 min @50Hz
- Fuse rating – T1AL, 250V
- Compressor weight up to 2.2 kg

Procurement Committee