

**CORRIGENDUM No: 4**

to the

**Bidding document for procurement of Goods**

**Ref No: ICB No: IOP/3-2015/NCE/1**

**Publication Ref: 2015/S 133-244623**

**Subject: Procurement of Capital equipment**

**Due to the changes imposed by the Final Beneficiary the Purchaser makes the following modifications to the Bidding Document:**

**• Original text of the Section VII. Schedule of Requirements, Technical Specification ANNEX II for Lot No. 5 Item 5.33 — Safety cabinet**

- 4 Independent alarm systems immediately alerts to any airflow obstruction
- 5 Motorized front window

**has been modified as follows:**

- 4 Independent alarm systems immediately alerts to any airflow obstruction
- 5 Interior dimension of cabinet: 1200x780x630mm (WxHxD)
- 6 Motorized front window

**• Original text of the Section VII. Schedule of Requirements, Technical Specification ANNEX II for Lot No. 5 Item 5.48 Laminar chamber - Biosafety cabinet:**

- 1.22.Vacuum tap G3/8 connection with long neck
- 2.0 Delivery period: 90days

**has been modified as follows:**

- 1.22.Vacuum tap G3/8 connection with long neck
- 2.0 Warranty period:24 month
- 3.0 Delivery period: 90days

Please find attached modified **Lot 5** of ANNEX 1 Technical part, Technical specifications .

**All other terms and conditions of the Bidding document remain unchanged.**

**The above amendments are integral part of the Bidding document.**

*Tranese Kostant*